STATE OF DELAWARE AMBULANCE INSPECTION SHEET

MISCELLANEOUS EQUIPMENT	YES	NO	N/A
1 Ambulance Cot with proper latching mechanisms (Minimum of 3 Patient straps- All strapes utilized as manufacture intended)			
1 folding stretcher or squad bench (optional)			
1 Stair Chair or Combination Stretcher/chair			
3 Blood Pressure Cuffs All Latex Free (large adult, adult and peds)			
1 Adult Stethoscope Latex Free			
1 Pediatric Stethoscope Latex Free			
2 Towels or similar disposable towels			
2 ~~1~~ Blanket			
12 triage Tags			
2 Liters of sterile normal saline solution or distilled water for irrigation (any size bottles totaling 2 liters)			
4 Ice Packs			
4 Hot Packs			
1 Burn kit w/sterile sheet			
2 working Pen Lights			
2 Emesis pans or basin			
1 Pillow Optional			
2 Sets of Sheets			
Glucometer with all required equipment			
2 Tubes Oral Glucose (not expired)			
Seat Belt Extender for at least 1 seat (unless seatbelt extends 60 inches)			
Pulse Ox			
Class 3 Breakaway Vest for each Crew member			

TOOLS	YES	NO	N/A
1 16oz claw hammer			
1 flat-head screw driver			
1 Phillips-head screw driver			
1 set of pliers			
1 adjustable wrench			
1 Wrecking bar 24" minimum			
1 Set of battery jumper Cables			
1 Spring-loaded center punch			

VEHICLE INSPECTION	YES	NO	N/A
Tires (acceptable tread) RF _____ RR _____ RIR _____ LF _____ LR _____ LIR _____			
Glass (no cracks or obstructions			
Environmental (operational Heating/Air Condition in crew and patient compartment)			
Mirror-operable(adjustable to offer field of vision)			
Wipers - operable (do they clear windshield)			
Horn - operable (manufacture installed)			
License plate (visible)			
Reflectors and Lenses(not cracked)			
Lights operable: turn signals,brake lights, head lights,tailights, parking lights, clearance lights			
Hood and latch, door latches(do they operate as designed			
Body Condition(holes,rust,major damage)			
Warning devices - operable: Emergency lights and sirens			
Two Way Radio Communication with reliable source			
Shock absorbers(excessive bouncing of unit)			
Exhaust System (leaks under vehicle or high CO reading in patient compartment)			

ALL ITEMS LISTED ON THIS FORM MUST OPERATE WITHIN ITS INTENDED FUNCTION

DO NOT SEND THIS FORM IN WITH THE APPLICATION FOR AMBULANCE PERMIT

*** CHILD CAR SEAT GUIDELINES:**

Convertible child safety seat with weight limits up to at least 65 ~~40~~ pounds and a 5 point harness. The child safety seat must meet all Federal Motor Vehicle Safety Seat Standard 213 requirements. The child safety seat must not exceed the disposal date recommended by the manufacturer. If the manufacturer does not specify, the car seat must not be older than 6 years old, as determined by the date of the manufacture sticker located on the restraint.

Date:_____ Time:_____

Member with unit at time of inspection :_____

Inspector signature:_____

Revisions 7/29/2013 - 10/15/13
Original 2/8/2012

STATE OF DELAWARE AMBULANCE INSPECTION SHEET

VENTILATION AND AIRWAY EQUIPMENT	YES	NO	N/A
1 Fixed oxygen inhalation system (min 500psi)			
2 size D or E portable oxygen bottles (min 500psi)			
1 Portable suction apparatus			
1 Fixed suction apparatus			
Bag Valve Mask (1 each Latex Free Adult-Child-Infant)			
Rigid and Flexible suction catheters (six different sizes from 5f to 14F, adult-child and infant)			
1 Set of Oropharyngeal airways (seven sizes including adult, child and infant)			
1 Set of Nasopharyngeal airways with water soluble lubricant (six sizes adult, child ~~and infant)~~			
4 Adult Non-Rebreathers - Latex Free			
4 Child Non-Rebreathers - Latex Free			
2 Infant Non-Rebreathers - Latex free (Optional)			
2 Adult Simple Mask Latex Free (Optional)			
2 Child Simple Mask Latex Free (Optional)			
4 Adult Nasal Cannulas - Latex Free			
4 Child Nasal Cannulas - Latex Free			
NuMask IOM (optional)			

Monitoring and Defibrillation	YES	NO	N/A
1 SAED (Semi-Automatic External Defibrillator)			
1 SAED Kit (1 Razor, 2 Adult / 1 Pediatric Pads or Key, 1 small towel, 1 pocket mask with a one-way valve)			

Immobilization Devices	YES	NO	N/A
Adult & Pediatric splints for 2 arms & 2 legs (inflatable splints, vacumn splints, SAM Splints, or wooden padded boards)			
1 Hare Traction splint or equivalent femur traction splint			
1 KED or equivalent extraction device			
3 back boards with appropriate securing straps attached to board (2 Adult & 1 Pediatric x-ray translucent) NO WOODEN BACKBOARDS			
3 complete head immobilization devices (Three Dimensional foam head blocks that adhere to spine board by Velcro, locking straps or click and lock head blocks)			
4 Adult & 4 Pediatric sized cervical stabilization devices/extrication collars (may be adjustable)			
1 Flexible stretcher (Reeves type style)			

OBSTETRICAL	YES	NO	N/A
2 O.B kits (sterile) to include - 1 drape sheet, 1 pair of rubber gloves, 1 receiving blanket, 1 bulb aspirator, 2 hand towels, 2 cord clamps or umbilical tape			

BANDAGES	YES	NO	N/A
1 Trauma Shears			
5 Adhesive Tape various sizes			
12 Combination Dressings (5"x9")			
24 Sterile Gauze Pads (4"x4")			
~~6~~ 8 Roll bandages, self adhering/kling various sizes ~~(3"x5 yards or 4"x5 yards)~~			
~~2 Roll bandages, self-adhering type/kling (6"x5 yards)~~			
2 Universal and/or multi-trauma dressings (10"x30")			
2 Elastic ACE (type) bandages			
3 Triangle bandages with minimum of 2 safety pins for each			
2 Latex Free Tourniquets Emergency Medical Tourniquets (such as C-A-T or M-A-T)			
2 packs of Hemostatic clotting ~~dressings~~ agent that is X-Ray Detectable			

Communications	YES	NO	N/A
Two way, reliable communication between EMT, Dispatcher, and Medical Direction (See SFPC Ambulance R ~~r~~egulation under Communications Requirements ~~Part X page 47 & 48~~)			

INFECTION CONTROL	YES	NO	N/A
2 Biohazard bags			
~~1 latex free kit to include - 2 pair each size latex free gloves, tourniquet, BVM , Nasal Cannula, and Non Rebreather Adult ped, & infant, BP Cuffs Lg Adult, adult and peds, Stethoscope adult and peds) ALL NEED TO BE LATEX FREE~~			
3 Hepa masks			
12 Eye/Mouth/Nose protection (disposable)			
3 disposable gowns			
12 pairs disposable gloves			
Disinfectant solution for cleaning equipment			
Disinfectant hand wash, commercial antimicrobial (towlette, spray or liquid)			

INJURY PREVENTION EQUIPMENT	YES	NO	N/A
1 Fire Extinguisher(minimum 5lb dry chemical) Inspected within last 2 years			
1 Flashlight			
1 Set of DOT triangle reflectors			
1 Sharps Container			
1 Child car seat (*see guideline note)			